



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 25176010

Part Number 48554
Exp Date 2026-09-25
Date of Manufacture 2025-06-25
UDI# (01)00385480485548(11)250625(17)260925(10)25176010

| | <u>Spec*</u> | <u>Mean Result</u> |
|------------------------|--------------|--------------------|
| Slope | ≥ 15.00 | 47.01 |
| pHox Ultra/CCX Ctl L4 | 1.7 - 2.3 | 1.8 |
| pHox Ultra/CCX Ctl L5 | 6.3 - 7.7 | 7.1 |
| Lin Soln Set A Level 1 | 14.0 - 20.0 | 18.9 |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

6/30/2025
Date