



CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 25188013

Ref Number 54652
Exp Date 2027-01-07
Date of Manufacture 2025-07-07
UDI# (01)00385480546522(11)250707(17)270107(10)25188013

	Ctl Level 1			8824157181		Ctl Level 2			8824157182
	<u>Spec</u>			<u>Pass/Fail</u>		<u>Spec</u>			<u>Pass/Fail</u>
HDL mg/dL	31	-	53	Pass	HDL mg/dL	68	-	102	Pass
TCHOL mg/dL	146	-	216	Pass	TCHOL mg/dL	294	-	410	Pass
TGL mg/dL	48	-	124	Pass	TGL mg/dL	290	-	430	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/18/2025
Date