nova biomedical

CERTIFICATE OF ANALYSIS

STP pHOx Ultra/CCX Lactate Membrane
Lot Number 25197011

Part Number 48554
Exp Date 2026-10-16
Date of Manufacture 2025-07-16
UDI# (01)00385480485548(11)250716(17)261016(10)25197011

	Spec*	Mean Result
Slope	≥ 15.00	34.65
pHOx Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHOx Ultra/CCX Ctl L5	6.3 - 7.7	7.1
Lin Soln Set A Level 1	14.0 - 20.0	17.9

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Z. Vegamwala
Approval

QC Inspector
Title

7/18/2025
Date