



CERTIFICATE OF ANALYSIS

Stat Profile pHox Plus L Calibrator Cartridge D
Lot Number 25190058

Ref Number 34088
Exp Date 2026-12-13

Date of Manufacture 2025-06-13

UDI# (01)00385480340885(11)250613(17)261213(10)25190058

Table with 3 columns: Cal-A L/N 25164046, Cal-B L/N 25164047, Cal-C L/N 25164048. Each column lists analytes (Na, HCO3, K, Cl, Ca, Glu, Lac) with their respective specifications and Pass/Fail status.

Table for Cal-D L/N 25164049 listing analytes pH @ 37C, Glu mg/dL, and Lac mmol/L with their specifications and Pass/Fail status.

Table for Reference L/N 25164045 listing analyte KCl mol/L with its specification and Pass/Fail status.

This certifies that this product was manufactured and tested at Nova Biomedical-Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical-Devices Quality Management Systems Requirements, Medical-Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/21/2025
Date