



**CERTIFICATE OF ANALYSIS**

STP pHox /Basic/Plus/L/C/M pCO2 Sensor  
Lot Number 25217018

Ref Number 21524

Exp Date 2027-08-05

Date of Manufacture 2025-08-05

UDI (01)00385480215244(11)250805(17)270805(10)25217018

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.50 - 11.30	Pass
Std A mV	0.00 +/- 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector  
Title

9/2/2025  
Date