



CERTIFICATE OF ANALYSIS

Nipro Glucose Control Solution Level 4
Lot Number 0425176304

Ref Number 95-094
Exp Date 2027.12.25

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	555 - 625	579

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "S. D. D.", written in black ink.

Approval

QC Inspector
Title

8/28/2025
Date