



CERTIFICATE OF ANALYSIS

Bioprofile Control Level 4
Lot Number 520640

Catalog Number 24617
Exp Date 2027-01

	<u>Spec</u>	<u>Pass/Fail</u>
Gln mmol/L	0.85 - 1.15	Pass
Glu mmol/L	0.40 - 0.60	Pass
Gluc g/L	0.85 - 1.15	Pass
Lac g/L	0.85 - 1.15	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Lac, Gluc.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

8/28/2025
Date