



CERTIFICATE OF ANALYSIS

Allegro UACR Test Cartridge
Lot Number 25198009

Ref Number 54214
Exp Date 2026-07-17
Date of Manufacture 2025-07-17
UDI# (01)00385480542142(11)250717(17)260717(10)25198009

	Ctl Level 1	Pass/Fail		Ctl Level 2	Pass/Fail
	Spec			Spec	
Albumin mg/L	29 - 53	Pass	Albumin mg/L	135 - 211	Pass
Creatinine mg/dL	60 - 104	Pass	Creatinine mg/dL	175 - 291	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

8/29/2025

Date