



**CERTIFICATE OF ANALYSIS**

STAT PROFILE® pHOx Ultra™/CCX™ pCO2 Sensor  
Lot Number 25191041

Ref Number 48600  
Exp Date 2027-07-10  
Date of Manufacture 2025-07-10  
UDI# (01)00385480486002(11)250710(17)270710(10)25191041

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.40 - 11.60	Pass
Std C mV	0.00 ± 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

8/21/2025  
Date