

OI-922 Page 54 of 65 Rev. BG Viald and Packaged Test Strip Release Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED
Catalog Number:	47940
Nova Biomedical P/N	47940
Target Range (mg/dL) Mean MPE (%) CV (%)	
60-75 ≤ 15 ≤ 6.5	PASS / FAIL
120-160 ≤ 15 ≤ 5.5	PASS / FAIL
250-290 ≤ 15 ≤ 5.0	PASS / FAIL
380-420 ≤ 15 ≤ 5.0	PASS / FAIL
500-540 ≤ 15 ≤ 5.0	PASS / FAIL
Combined MPE (%) ≤ 10%	PASS / FAIL
Qty:	304
Lot No.	2122200309
Mfg. Date:	2022.07.19
Expiration Date:	2025.01.31

M. B. Bue
Quality Control Manager

7/27/2022
Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

OI-922 Page 55 of 65 Rev. BG Vial and Packaged Test Strip Release Test Strips

nova
biomedical

ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2

RELEASE TEST RESULTS

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	47940
Nova Biomedical P/N	47940
Lot No.	2122200309

(Spiked whole blood) Target Range:60-75			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤6.5		≤15
68	3.7	69.1	3.5

(Spiked whole blood) Target Range:120-160			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.5		≤15
150	4.1	57.0	4.8

(Spiked whole blood) Target Range:250-290			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
276	3.8	287.0	4.4

(Spiked whole blood) Target Range:380-420			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
393	2.9	411.5	4.7

(Spiked whole blood) Target Range:500-540			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
503	3.2	520.8	3.8

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454