



**CERTIFICATE OF ANALYSIS**

BioProfile FLEX 2 pH/Gas Reference Sensor  
Lot Number 25225019

Part Number 57467  
Exp Date 2027-02-13  
Date of Manufacture 2025-08-13

|                | <u>Spec*</u> | <u>Pass/Fail</u> |
|----------------|--------------|------------------|
| pH Slope       | 8.70 - 11.80 | Pass             |
| BioPro Ctl L 1 | 6.628-6.728  | Pass             |
| BioPro Ctl L 3 | 7.322-7.422  | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

\* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

08/19/2025

Date