



CERTIFICATE OF ANALYSIS

BioPro Flex2 ESM Reagent Cartridge
Lot Number 25211009

Ref Number 56227
Exp Date 2026-07-16

Solution S L/N 25197049

	<u>Spec</u>	<u>Pass/Fail</u>
pH	6.900 - 7.100	Pass

Solution DP L/N 25197050

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl %	0.43 - 0.73	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/4/2025

Date