



**CERTIFICATE OF ANALYSIS**

Allegro Lipids Test Cartridge  
Lot Number 25230009

Ref Number 54652  
Exp Date 2027-02-18  
Date of Manufacture 2025-08-18  
UDI# (01)00385480546522(11)250818(17)270218(10)25230009

	Ctl Level 1				Ctl Level 2			
	<u>Spec</u>		<u>Pass/Fail</u>		<u>Spec</u>		<u>Pass/Fail</u>	
HDL mg/dL	31	- 53	Pass	HDL mg/dL	68	- 102	Pass	
TCHOL mg/dL	146	- 216	Pass	TCHOL mg/dL	294	- 410	Pass	
TGL mg/dL	48	- 124	Pass	TGL mg/dL	290	- 430	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

9/4/2025  
Date