



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ Lactate Sensor  
Lot Number 25204025

Part Number 48553  
Exp Date 2027-07-23

Date of Manufacture 2025-07-23

UDI# (01)00385480485531(11)250723(17)270723(10)25204025

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	78 - 92	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Z. V. [Signature]*  
Approval

QC Inspector  
Title

8/6/2025  
Date