



CERTIFICATE OF ANALYSIS

STP pHox Plus Chloride Sensor
Lot Number 25169036

Part Number 27812
Exp Date 2026-09-18
Date of Manufacture 2025-06-18
UDI# (01)10385480278123(11)250618(17)260918(10)25169036

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.00	Pass
Std A mV	55.00 - 95.00	Pass
pHOx Ctl L1	120.0 - 129.0	Pass
pHOx Ctl L3	79.6 - 89.6	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/10/2025
Date