nova biomedical

CERTIFICATE OF ANALYSIS

Bio pHOx Cal Pack — Part Number 41874
Lot Number 25246013 — Exp Date 2027-02

Std A L/N 25232018

	Spec	Pass/Fail
Na mmol/L	139.0 - 141.0	Pass
HCO3 mmol/L	19.6 - 20.4	Pass

Std B L/N 25232024

	Spec	Pass/Fail
HCO3 mmol/L	39.2 - 40.8	Pass

Std C L/N 25232006

	Spec	Pass/Fail
pH @ 37C	7.352 - 7.358	Pass

Std D L/N 25232019

	Spec	Pass/Fail
pH @ 37C	6.802 - 6.808	Pass
Na mmol/L	69.5 - 70.5	Pass

Reference L/N 25231049

	Spec	Pass/Fail
KCl mol/L	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCl.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

9/10/2025
Date