



CERTIFICATE OF ANALYSIS

STP Prime+ Vet Cal-Cartridge 200 Sample w. Creat/BUN
Lot Number 25246020

Ref Number 58405
Exp Date 2026-12-13
Date of Manufacture 2025-06-13

Table with 3 columns: Cal A L/N 25168010, Cal B L/N 25168011, Cal C L/N 25168012. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN, Creat and tHb g/dL.

Table with 3 columns: Cal D L/N 25168013, Flush L/N 25232007, Reference L/N 25164044. Rows include pH @ 37C and PO2.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval [Signature]

QC Inspector
Title

9/10/2025
Date