



**CERTIFICATE OF ANALYSIS**

STP pHox/pHOx Basic/BiopHOx Air Detector  
Lot Number 25217026

Ref Number 21513  
Date of Manufacture 2025-08-05

	<u>Spec</u>	<u>Pass/Fail</u>
AD1 mV	1400 - 2600	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to be "J. M. ...", written over a faint red horizontal line.

Approval

QC Inspector

Title

8/14/2025

Date