



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ Potassium Sensor
Lot Number 25234035

Ref Number 48582

Exp Date 2026-11-22

Date of Manufacture 2025-08-22

UDI# (01)00385480485821(11)250822(17)261122(10)25234035

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope | 9.00 to 11.00 | Pass |
| Std C mV | 0 ± 50.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/11/2025

Date