



CERTIFICATE OF ANALYSIS

Allegro HbA1c Test Cartridge
Lot Number 25183006

Ref Number 54215
Exp Date 2026-07-02
Date of Manufacture 2025-07-02
UDI# (01)00385480542159(11)250702(17)260702(10)25183006

Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HbA1c %	4.3 - 6.1	Pass	HbA1c %	7.1 - 10.7	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/2/2025

Date