



CERTIFICATE OF ANALYSIS

StatSensor Creatinine Control Solution Level 2
Lot Number 5025224242

Ref Number 43922
Exp Date 2027-08-12
Date of Manufacture 2025-08-12
UDI# (01)00385480439220(11)250812(17)270812(10)5025224242

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	1.95 - 2.25	2.06

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

8/29/2025
Date