



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 100 Sample
Lot Number 25237040

Ref Number 57825
Exp Date 2027-01-31

Date of Manufacture 2025-07-31

UDI# (01)00385480578257(11)250731(17)270131(10)25237040

Table with 3 columns: Cal A L/N 25233007, Cal B L/N 25213015, Cal C L/N 25216028. Rows include analytes like HCO3, Na, K, Cl, Ca, Mg, Glu, Lac with Spec and Pass/Fail columns. Includes a row for tHb g/dL with Assigned Value* and Result.

Table with 3 columns: Cal D L/N 25213017, Flush L/N 25213012, Reference L/N 25212038. Rows include pH @ 37C, PO2, KCl with Spec and Pass/Fail columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Signature: Romanus Ahtu
Approval
QC Inspector
Title
9/3/2025
Date