



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 200 Sample
Lot Number 25251018

Ref Number 57826
Exp Date 2027-01-01

Date of Manufacture 2025-07-01

UDI# (01)00385480578264(11)250701(17)270101(10)25251018

Table with 3 columns: Cal A L/N 25182076, Cal B L/N 25182077, Cal C L/N 25182078. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, and tHb g/dL with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25182079, Flush L/N 25241026, Reference L/N 25182080. Rows include pH @ 37C, PO2, and KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

9/19/2025
Date