



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pH Sensor
Lot Number 25238001

Ref Number 21522
Exp Date 2027-08-26
Date of Manufacture 2025-08-26
UDI# (01)00385480215220(11)250826(17)270826(10)25238001

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----------|-----------------|------------------|
| Slope | 10.0 - 10.8 | Pass |
| Std A mV | 0.00 +/- 110.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Z. Vagstad
Approval

QC Inspector
Title

9/18/2025
Date