



CERTIFICATE OF ANALYSIS

NOVA PRO Glucose and Ketone Control Solution Level 2
Lot Number 7525245302

Ref Number 52332
Exp Date 2028-03-02
Date of Manufacture 2025-09-02
UDI# (01)00385480523325(11)250902(17)280302(10)7525245302

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	98 - 128	119
Ketone mmol/L	0.6 - 0.8	0.8

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

9/18/2025
Date