



**CERTIFICATE OF ANALYSIS**

STP Prime Plus Vet Reference Cartridge  
Lot Number 25255048

Ref Number 59345  
Exp Date 2027-03-12  
Date of Manufacture 2025-09-12

Internal QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	137.2 - 145.2	Pass
K mmol/L	3.67 - 4.17	Pass
CL mmol/L	122.5 - 131.5	Pass

  

External QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	134.3 - 142.3	Pass
K mmol/L	3.73 - 4.23	Pass
CL mmol/L	118.9 - 127.9	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/18/2025

Date