



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 400 Sample
Lot Number 25258045

Ref Number 57828
Exp Date 2027-02-26

Date of Manufacture 2025-08-26

UDI# (01)00385480578288(11)250826(17)270226(10)25258045

Table with 3 columns: Cal A L/N 25239015, Cal B L/N 25239016, Cal C L/N 25239017. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, and tHb g/dL with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25239018, Flush L/N 25251004, Reference L/N 25238063. Rows include pH @ 37C, PO2, and KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Signature of Luciano Acheta

Approval

QC Inspector
Title

9/23/2025
Date