



**CERTIFICATE OF ANALYSIS**

STP Prime ABG Cal Cartridge 200 Sample  
Lot Number 25202011

Ref Number 53360  
Exp Date 2027-01-03  
Date of Manufacture 2025-07-03  
UDI# (01)00385480533607(11)250703(17)270103(10)25202011

Cal A	L/N	25184016	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.353	N/A
	pCO2 mmHg	25.2	N/A

Cal B	L/N	25184017	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	25184019	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	51.4	N/A
	pO2 mmHg	102.7	N/A

Reference	L/N	25184018	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

08/27/2025

Date