



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 300 Sample w. Creat/BUN
Lot Number 25239047

Ref Number 57833
Exp Date 2027-02-20

Date of Manufacture 2025-08-20

UDI# (01)00385480578332(11)250820(17)270220(10)25239047

Table with 3 columns: Cal A L/N 25240016, Cal B L/N 25233008, Cal C L/N 25233009. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, Creat mg/dL, pH, tHb g/dL, and Assigned Value\*/Result.

Table with 3 columns: Cal D L/N 25233010, Flush L/N 25238035, Reference L/N 25233006. Rows include pH @ 37C, PO2, and KCl.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Signature of Luciano Acheta

Approval

QC Inspector
Title

9/24/2025
Date