



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 400 Sample w. Creat/BUN
Lot Number 25239047

Ref Number 57834
Exp Date 2027-02-20

Date of Manufacture 2025-08-20

UDI# (01)00385480578349(11)250820(17)270220(10)25239047

Table with 3 columns: Cal A L/N 25240016, Cal B L/N 25233008, Cal C L/N 25233009. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN, Creat with Spec and Pass/Fail columns.

Table with 3 columns: Cal D L/N 25233010, Flush L/N 25238035, Reference L/N 25233006. Rows include pH @ 37C, PO2, KCl with Spec and Pass/Fail columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Signature: Luciano Aichta

Approval

QC Inspector
Title

9/24/2025
Date