



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 100 Sample
Lot Number 25258045

Ref Number 57825
Exp Date 2027-02-26

Date of Manufacture 2025-08-26

UDI# (01)00385480578257(11)250826(17)270226(10)25258045

Table with 3 columns: Cal A L/N 25239015, Cal B L/N 25239016, Cal C L/N 25239017. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac mmol/L and tHb g/dL with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25239018, Flush L/N 25251004, Reference L/N 25238063. Rows include pH @ 37C, PO2, KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Signature: Romanus Ahtu
Approval
QC Inspector
Title
9/26/2025
Date