



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 700 Sample
Lot Number 25251016

Ref Number 65141

Exp Date 2026-11-20

Date of Manufacture 2025-05-20

UDI# (01)00385480651417(11)250520(17)261120(10)25251016

Table with 3 columns: Cal A L/N 25140079, Cal B L/N 25140080, Cal C L/N 25140081. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, pH, tHb g/dL, and Assigned Value\*/Result.

Table with 3 columns: Cal D L/N 25140082, Flush L/N 25241026, Reference L/N 25140098. Rows include pH @ 37C, PO2, and KCl with Spec and Pass/Fail columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Signature of P. Hernandez

Approval

QC Inspector

Title

9/24/2025

Date