



CERTIFICATE OF ANALYSIS

Nova Lactate Plus Control Solution Level 1
Lot Number 3825226301

Ref Number 40815
Exp Date 2028-02-14
Date of Manufacture 2025-08-14
UDI# (01)00385480408158(11)250814(17)280214(10)3825226301

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	1.3 - 1.5	1.3

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/24/2025

Date