



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pO2 Sensor
Lot Number 25246047

Ref Number 21521
Exp Date 2027-09-03

Date of Manufacture 2025-09-03

UDI# (01)00385480215213(11)250903(17)270903(10)25246047

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------|----------------|------------------|
| Slope | -3.00 to -6.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

9/29/2025
Date