



CERTIFICATE OF ANALYSIS

StatSensor Creatinine Control Solution Level 3
Lot Number 5025212243

Ref Number 43923

Exp Date 2027-07-31

Date of Manufacture 2025-07-31

UDI# (01)00385480439237(11)250731(17)270731(10)5025212243

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	5.50 - 7.00	5.98

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

8/25/2025

Date