



CERTIFICATE OF ANALYSIS

STP pHox Plus Reference Sensor
Lot Number 25260016

Ref Number 28954
Exp Date 2027-03-17
Date of Manufacture 2025-09-17
UDI# (01)00385480289542(11)250917(17)270317(10)25260016

	<u>Spec</u>	<u>Pass/Fail</u>
Na Slope	9.00 - 11.00	Pass
ADT 4	≥ 220.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Z. Verdamwald

Approval

QC Inspector
Title

10/2/2025
Date