



CERTIFICATE OF ANALYSIS

Prime ES Diluent Solution
Lot Number 25216027

Ref Number 56452
Exp Date 2026-08-04
Date of Manufacture 2025-08-04
UDI# 00385480564526(11)250804(17)260804(10)25216027

	Spec	Pass/Fail
pH	6.00-8.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/6/2025
Date