nova biomedical

CERTIFICATE OF ANALYSIS

STP pHOx Plus Sodium Sensor
Lot Number 25259033

Ref Number 27810
Exp Date 2027-03-16
Date of Manufacture 2025-09-16
UDI# (01)00385480278102(11)250916(17)270316(10)25259033

	Spec	Pass/Fail
Slope	9.00 - 10.50	Pass
Std C mV	-80.00 - +80.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

9/30/2025
Date