



CERTIFICATE OF ANALYSIS

STP pHox Plus K Sensor
Lot Number 25247010

Part Number 27811
Exp Date 2026-12-04

Date of Manufacture 2025-09-04

UDI# (01)00385480278119(11)250904(17)261204(10)25247010

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.60	Pass
Std A mV	0 ± 60.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/02/2025
Date