



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 CDV Reagent Pack
Lot Number 25272021

Ref Number 55297
Exp Date 2026-08-22
Date of Manufacture 2025-08-22

	CDV Bottle Pack L/N:	25272020	Exp Date:	2026-08-22		
Soln SF L/N	25234021		Soln DP L/N	25234022		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
pH	6.900 - 7.100	Pass	Depro Soln	N/A	Present	
Cal C L/N	25234020		Soln T L/N	25241030		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
3% Detergent	N/A	Present	0.4% Trypan	N/A	Present	

CDV Reagent Cartridge L/N:	25272018	Exp Date:	2026-11-22
Soln PF L/N	25234027		
	<u>Spec</u>	<u>Pass/Fail</u>	
pH	6.900 - 7.100	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/8/2025
Date