



**CERTIFICATE OF ANALYSIS**  
**StatStrip Glucose Control Solution Level 3**

Nova Ref Number 54522  
LifeScan Ref Number 02315602  
Lot Number 0425241243  
Date of Manufacture 2025-08-29  
Exp Date 2027-08-29  
UDI # (01)00385480545228(11)250829(17)270829(10)0425241243

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	299

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Awishi M*

Approval

QC Inspector

Title

10/09/2025

Date