



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 100 Sample
Lot Number 25269056

Ref Number 57825
Exp Date 2027-03-12

Date of Manufacture 2025-09-12

UDI# (01)00385480578257(11)250912(17)270312(10)25269056

Table with 3 columns: Cal A L/N 25255039, Cal B L/N 25255040, Cal C L/N 25255041. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac mmol/L and tHb g/dL with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25255042, Flush L/N 25258025, Reference L/N 25255025. Rows include pH @ 37C, PO2, KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/21/2025
Date