



CERTIFICATE OF ANALYSIS

Stat Profile pHox/Basic Calibrator Cartridge D  
Lot Number 25274036

Ref Number 23934  
Exp Date 2027-03-24  
Date of Manufacture 2025-09-24  
UDI# (01)00385480239349(11)250924(17)270324(10)25274036

Cal A L/N 25268023

Cal B L/N 25269027

Cal C L/N 25267006

	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HCO3 mmol/L	20.1 - 20.9	Pass	HCO3 mmol/L	40.2 - 41.8	Pass	pH @ 37C	7.342 - 7.348	Pass
Na mmol/L	139.3 - 141.3	Pass						

Cal D L/N 25267007

Reference L/N 25267004

	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
pH @ 37C	6.792 - 6.798	Pass	KCl mol/L	1.8 - 2.2	Pass
Na mmol/L	71.5 - 72.5	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

10/15/2025  
Date