



CERTIFICATE OF ANALYSIS

Bio pHox Calibrator Cartridge
Lot Number 25274035

Ref Number 41874
Exp Date 2027-03

Cal A L/N 25268023

Cal B L/N 25269027

Cal C L/N 25267006

	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HCO3 mmol/L	139.0 - 141.0	Pass	HCO3 mmol/L	39.2 - 40.8	Pass	pH @ 37C	7.352 - 7.358	Pass
Na mmol/L	19.6 - 20.4	Pass						

Cal D L/N 25267007

Reference L/N 25267004

	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
pH @ 37C	6.802 - 6.808	Pass	KCl mol/L	1.8 - 2.2	Pass
Na mmol/L	69.5 - 70.5	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/15/2025
Date