



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 200 Sample w. Creat/BUN
Lot Number 25279029

Ref Number 57832
Exp Date 2026-05-08

Date of Manufacture 2024-11-08

UDI# (01)00385480578325(11)241108(17)260508(10)25279029

Table with 3 columns: Cal A L/N 24313057, Cal B L/N 24313058, Cal C L/N 24313059. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN, Creat, and tHb with Spec and Pass/Fail columns.

Table with 3 columns: Cal D L/N 24313060, Flush L/N 25262033, Reference L/N 24313020. Rows include pH @ 37C, PO2, and KCl with Spec and Pass/Fail columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

10/22/2025
Date