



CERTIFICATE OF ANALYSIS

Allegro CRP Test Cartridge
Lot Number 25203094

Ref Number 62050
Exp Date 2026-07-22

Date of Manufacture 2025-07-22

UDI# (01)00385480620505(11)250722(17)260722(10)25203094

	Ctl Level 1			Ctl Level 2	
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
CRP mg/L	10.6 - 24.6	Pass	CRP mg/L	50.0 - 96.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

08/14/2025
Date