nova biomedical

CERTIFICATE OF ANALYSIS

BioProfile CDV Reagent Pack
Lot Number 25213035

Ref Number 48796
Exp Date 2026-06

CDV Bottle Pack L/N 25213034 Exp Date 2026-06

C Soln L/N 25164038

	Spec	Result
3% Detergent	N/A	Present
pH	> 9.0	Pass

T Soln L/N 25164039

	Spec	Result
0.4% Trypan	N/A	Present
Particle Count	≤ 2.0	Pass

CDV Cartridge L/N 25213033 Exp Date 2026-06

S Soln L/N 25167056

	Spec	Pass/Fail
pH	6.900 - 7.100	Pass

D Soln L/N 25164040

	Spec	Result
%NaOCl	0.43 - 0.73	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

8/20/2025
Date