



**CERTIFICATE OF ANALYSIS**

Nova Primary Calibrator Cartridge 150 Sample  
Lot Number 25177020

Ref Number 63167  
Exp Date 2026-12-23  
Date of Manufacture 2025-06-23  
UDI# (01)00385480631679(11)250623(17)261223(10)25177020

Cal A L/N 25174031			Diluent L/N 25176011		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Glu mg/dL	19.9 - 20.1	Pass	pH	4.98 - 5.02	Pass
			Buffer	N/A	Present

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Inspector

Title

8/12/2025

Date