



CERTIFICATE OF ANALYSIS

Allegro UACR Test Cartridge
Lot Number 25288004

Ref Number 54214

Exp Date 2026-10-15

Date of Manufacture 2025-10-15

UDI# (01)00385480542142(11)251015(17)261015(10)25288004

| | Ctl Level 1 | | | Ctl Level 2 | |
|------------------|-------------|-----------|------------------|-------------|-----------|
| | Spec | Pass/Fail | | Spec | Pass/Fail |
| Albumin mg/L | 36 - 60 | Pass | Albumin mg/L | 135 - 211 | Pass |
| Creatinine mg/dL | 61 - 105 | Pass | Creatinine mg/dL | 175 - 291 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/24/2025
Date