



CERTIFICATE OF ANALYSIS

Nova Primary Glucose Controls Levels 1 and 2

Ref Number 63164

Lot Number 25287012

Exp Date 2027-03-18

Date of Manufacture 2025-09-18

UDI# (01)10385480631645(11)250918(17)270318(10)25287012

Level 1			Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Glu mg/dL	71 - 73	Pass	Glu mg/dL	188 - 196	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

10/27/2025

Date